**Product** Launch serial number M unit serial number S unit serial number Test date

XK10 system 4PV151 4PV153 4PV152 16 May 2024



## **Calibration certificate**

Straightness performance		
Accuracy*	±0.01A±1 μm	
Resolution	0.1 μm	

Squareness performance			
Accuracy**	±0.01A/M ±2/M ±10 μm/m		
Accuracy***	±0.01A/M ±2/M ±4 μm/m		
Resolution	0.1 μm		

Flatness performance			
Accuracy****	±0.01A±1±(1+1.1M) μm		
Resolution	0.1 μm		

<sup>\*</sup> Provided measurement is within ±500µm of PSD centre

M = Axis measurement distance (m)

A = Maximum absolute PSD deviation during measurement (µm)

Accuracy values are reported to a statistical confidence of 95% (k=2).

We certify that this instrument has been tested and complies with our published standards and specifications.

Authorised signature	Signatory	Position	Issue date
	Dave Wall	Director and General Manager	10 June 2024

This certificate may not be reproduced other than in full, except with the prior written approval of:

Renishaw plc Laser & Calibration Products Division Bath Road, Woodchester Stroud Gloucestershire

GL5 5EY United Kingdom

<sup>\*\*</sup> Without squareness calibration factor

<sup>\*\*\*</sup> With squareness calibration factor

<sup>\*\*\*\*</sup> Within a 90° sector of launch unit rotation