

Product  
Serial number  
Date of calibration

HS20 laser  
30G135  
22nd Aug 2024


**RENISHAW**  
apply innovation™



# Calibration certificate

**Specification**                      *Vacuum wavelength*                      0.6329905770  $\mu\text{m} \pm 0.1$  ppm  
*Equivalent frequency*                      473612829.2 MHz

### Measured values and uncertainties of calibration



Results	Value (MHz)	Value (ppm)
Laser frequency:	473612830.4	-
Laser frequency error:	1.2	0.003
Stability (peak-to-peak):	3.1	0.007
Maximum laser frequency error:	2.8	0.006
Uncertainty of measurement (k=2):	$\pm 5.9$	$\pm 0.01$

Reference standards	Ref. no.	Lab	Certificate no.	Calibration date
Iodine stabilised HeNe laser	RUK27030	NPL	2022080011-LL03	11th Aug 2022
Frequency counter	RUK18299	UKAS0152	U380646	13th Nov 2023
Reference HeNe laser	XL-80 REF29	Renishaw	2AH074-240814-00	14th Aug 2024
Test procedure	WI-10840			

**Laser measurement system accuracy:** Based on this calibration, when this HS20 laser is used with a Renishaw RCU10 compensator and a Renishaw air temperature sensor (*both within specification*) the laser measurement system accuracy (k=2) in linear measurement mode will be within:  $\pm 1.0$  ppm (see the system manual for details).

Authorised signature	Signatory	Position	Issue date
	Dave Wall	Director & General Manager	30th Aug 2024

*This certificate may not be reproduced other than in full, except with the prior written approval of:*

Renishaw plc  
Laser & Calibration Products Division  
Bath Road, Woodchester  
Stroud  
Gloucestershire  
GL5 5EY  
United Kingdom  
Tel +44 (0) 1453 524524

**Certificate number**  
**30G135-240822-00**

L-9908-0888/04

# Calibration notes

1. **Lasers (XM, XL, ML, HS and RLU)** are calibrated by comparison to a reference HeNe laser using an optical beat frequency technique. Reference lasers are routinely calibrated against an iodine-stabilised HeNe laser supplied by the National Physical Laboratory (NPL), or by a national standards laboratory. All frequency measurements are taken over a 1 hour period.
2. **Air pressure and relative humidity (RH) sensors are installed in a compensator (XC and RCU).**  
The air pressure sensors are calibrated over 650 mbar to 1150 mbar range in a temperature controlled oven by direct comparison with a reference pressure meter. The RH sensors (where fitted) are certified by the manufacturer to be within specification. They are calibrated by comparison of the readings with those from a reference RH meter at a single applied humidity.
3. **Air and material temperature sensors (XC and RCU)** are calibrated by direct comparison with transfer platinum resistance thermometers (PRTs) in a temperature controlled water bath over 0 °C to 40 °C (50 °C for material sensor). The transfer PRTs are routinely calibrated against reference PRTs.
4. **Rotary axis calibrators (XR20)** are calibrated using a HeNe laser angular interferometer.
5. **Ballbar transducers (QC20 and QC10)** are calibrated using a HeNe laser interferometer. The scale factor (QC10 only) is calculated and must be entered into the Renishaw application software prior to use.
6. **Ballbar calibrators** are calibrated by direct comparison with a reference ballbar calibrator (calibrated by a national standards laboratory) using a reference ballbar as a transfer standard. The measured values for the ballbar calibrator must be entered into the Renishaw application software prior to use.
7. **Traceability.** All the reference standards (listed overleaf) used in these calibrations are traceable either directly to major international metrology institutes who have signed the CIPM Mutual Recognition Agreement (e.g. NPL: UK; LNE: France; NIST: USA; PTB: Germany; NMIJ: Japan) or to a national accreditation body (e.g. UKAS: UK; A2LA: USA).
8. **Environment.** The equipment used for calibration is in a facility held between 15 °C and 25 °C.
9. **Uncertainty calculations.** The uncertainty calculations have been carried out according to the European Co-operation for Accreditation document EA-4/02.
10. **Quality accreditation.** All calibrations above are covered by Renishaw's ISO 9001 quality assurance system. The system is audited and certified by an accredited agency.
11. **Re-calibration.** Customers may wish to confirm that systems are performing within published specifications over time. If so, it is recommended that they should be periodically re-calibrated. Please note that compensators and temperature sensors are re-calibrated only at a single applied temperature, air pressure and humidity. Please refer to the appropriate system manual for further details.