







Product
Launch serial number
M unit serial number
S unit serial number
Calibration date

XK10 system
4PV151
4PV153
4PV152
16 May 2024

Calibration certificate



Results	M unit		S unit	
Scale error (%)	V: 0.0082	H: 0.0082	V: 0.0055	H: 0.0000
Uncertainty of measurement (k=2):	±0.48%			




Results	North direction	South direction	East direction	West direction
Flatness (µm)	1.00	6.00	4.00	1.00
Uncertainty of measurement (k=2):	±1 µm			
Squareness Factor (mm/m):	-0.0005			

Reference standards	Ref. no.	Lab	Certificate no.	Calibration date
Linear gauge	INV-0214	Mitutoyo	20230458	19 October 2023
Calibrated artefact	INV-0288	Easy-Laser	24050871	08 May 2024
Calibrated artefact	INV-0289	Easy-Laser	24050815	08 May 2024
Test procedure	WI-100125			

Calibration notes

XK10 transceivers are calibrated by comparison to a linear gauge.

Authorised signature	Signatory	Position	Issue date
	Dave Wall	Director and General Manager	10 June 2024

This certificate may not be reproduced other than in full, except with the prior written approval of:

Renishaw plc
Laser & Calibration Products Division
Bath Road, Woodchester
Stroud
Gloucestershire
GL5 5EY
United Kingdom
Tel. +44 (0) 1453 524524

Certificate number 4PV153-160524-00
