

UK DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer, Renishaw plc. The object of the declaration is identified below: -

Product Name: Virsa Raman Analyser
Description: Portable, rack-mountable, fibre-launched laser, Raman spectrographic analyser
Part no. 0
Serial no. 0

The object of the declaration described above is in conformity with all relevant UK Statutory Instruments (and their amendments): -

S.I. 2016 No. 1101 The Electrical Equipment (Safety) Regulations 2016
S.I. 2016 No. 1091 Electromagnetic Compatibility Regulations 2016
S.I. 2012 No. 3032 The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 - RoHS

and complies with the following technical standards: -

EN ISO 12100:2010 Safety of machinery - General principles for design - Risk assessment and risk reduction
EN 61010-1:2010+ A1:2019 Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements
EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
EN 60825-1:2014 Safety of laser products - Part 1: Equipment classification and requirements
EN IEC 63000:2018 Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

The person authorised to sign for and on behalf of the manufacturer Renishaw Plc.:

Name Tim Mills
Job Title Senior Project & Quality Manager, Spectroscopy Division

Signature 

Dated 5th May 2022

Place Renishaw plc, New Mills, Wotton-under-Edge, Gloucestershire, GL12 8JR, UK